

SUPERVISOR TEST

WASHINGTON COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106406

Date: 10/18/2021

OPERATOR NAME:  
WILLIAM M GREENE  
Permit #: 48760  
Operator Agency/Dept: WASHINGTON CO. SO  
Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:56
INTERNAL STANDARD	VERIFIED	19:56
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	19:57
BLANK TEST	0.00	19:57
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	19:58
BLANK TEST	0.00	19:59
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	19:59
BLANK TEST	0.00	20:00

# 647  
*[Handwritten Signature]*

