

SUPERVISOR TEST

 WASHINGTON CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 100284

Date: 10/25/2021

OPERATOR NAME:
 MICHAEL O WEBER
 Permit #: 46067
 Operator Agency/Dept: WASHINGTON CO. SO
 Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:46
INTERNAL STANDARD	VERIFIED	06:46
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:46
BLANK TEST	0.00	06:47
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	06:48
BLANK TEST	0.00	06:48
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:49
BLANK TEST	0.00	06:50

