

SOLUTION CHANGE

 WASHINGTON CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 100284

Date: 11/01/2021

OPERATOR NAME:
 MATTHEW K JACKSON
 Permit #: 40078
 Operator Agency/Dept: WASHINGTON CO. SO
 Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:16
INTERNAL STANDARD	VERIFIED	07:16
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:16
BLANK TEST	0.00	07:17
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:17
BLANK TEST	0.00	07:18
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:18
BLANK TEST	0.00	07:19

