

SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 10/25/2021

OPERATOR NAME:
RONALD D TAYLOR III
Permit #: 49003
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 21280 --

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:43
INTERNAL STANDARD	VERIFIED	04:43
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	04:43
BLANK TEST	0.00	04:44
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	04:44
BLANK TEST	0.00	04:45
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	04:46
BLANK TEST	0.00	04:46

