

SOLUTION CHANGE

-----  
GRANVILLE VILLAGE  
POLICE DEPT  
DATAMASTER dmt: 113306  
-----

Date: 11/02/2021

OPERATOR NAME:  
THOMAS J ZOISTOSKI  
Permit #: 36771  
Operator Agency/Dept: GRANVILLE VILLAGE PD  
Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:51
INTERNAL STANDARD	VERIFIED	11:51
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	11:51
BLANK TEST	0.00	11:52
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	11:52
BLANK TEST	0.00	11:53
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	11:53
BLANK TEST	0.00	11:54

