

SOLUTION CHANGE

CAMBRIDGE VILLAGE  
POLICE DEPT  
DATAMASTER dmt: 126706

Date: 11/01/2021

OPERATOR NAME:  
ROBERT M DANKO  
Permit #: 39105  
Operator Agency/Dept: CAMBRIDGE VILL. PD  
Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:11
INTERNAL STANDARD	VERIFIED	14:11
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:11
BLANK TEST	0.00	14:12
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:12
BLANK TEST	0.00	14:13
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:13
BLANK TEST	0.00	14:14

