

SUPERVISOR TEST

HUDSON FALLS  
VILLAGE POLICE DEPT.  
DATAMASTER dmt: 102408

Date: 10/25/2021

OPERATOR NAME:  
DAVID M COSTELLO  
Permit #: 40891  
Operator Agency/Dept: HUDSON FALLS VIL. PD  
Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:52
INTERNAL STANDARD	VERIFIED	21:52
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	21:52
BLANK TEST	0.00	21:53
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	21:53
BLANK TEST	0.00	21:54
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	21:54
BLANK TEST	0.00	21:55

