

SUPERVISOR TEST

WHITEHALL VILLAGE  
POLICE DEPT.  
DATAMASTER dnt: 113806

Date: 10/23/2021

OPERATOR NAME:  
BRYAN J GRECO  
Permit #: 44077  
Operator Agency/Dept: WHITEHALL, VILL. P.D.  
Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:45
INTERNAL STANDARD	VERIFIED	22:45
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	22:45
BLANK TEST	0.00	22:46
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	22:46
BLANK TEST	0.00	22:47
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	22:48
BLANK TEST	0.00	22:48

