

SUPERVISOR TEST

WHITEHALL VILLAGE
POLICE DEPT.
DATAMASTER dmt: 113806

Date: 10/10/2021

OPERATOR NAME:
DAVID F BUXTON
Permit #: 33398
Operator Agency/Dept: WHITEHALL VILL. P.D.
Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:43
INTERNAL STANDARD	VERIFIED	16:43
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	16:44
BLANK TEST	0.00	16:45
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	16:45
BLANK TEST	0.00	16:46
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	16:46
BLANK TEST	0.00	16:47

