

SUPERVISOR TEST

-----  
CAMBRIDGE VILLAGE  
POLICE DEPT  
DATAMASTER dmt: 126706  
-----

Date: 11/15/2021

OPERATOR NAME:  
ROBERT M DANKO  
Permit #: 39105  
Operator Agency/Dept: CAMBRIDGE VILL. PD  
Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:05
INTERNAL STANDARD	VERIFIED	07:05
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:05
BLANK TEST	0.00	07:06
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:06
BLANK TEST	0.00	07:07
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:08
BLANK TEST	0.00	07:08

