

SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 11/15/2021

OPERATOR NAME:
THOMAS J ZOVISTOSKI
Permit #: 36771
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:40
INTERNAL STANDARD	VERIFIED	05:41
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	05:41
BLANK TEST	0.00	05:42
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	05:42
BLANK TEST	0.00	05:43
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	05:43
BLANK TEST	0.00	05:44

