

SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 11/23/2021

OPERATOR NAME:
THOMAS J ZOVVISTOSKI
Permit #: 36771
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:40
INTERNAL STANDARD	VERIFIED	06:40
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	06:41
BLANK TEST	0.00	06:41
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	06:42
BLANK TEST	0.00	06:43
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	06:43
BLANK TEST	0.00	06:44

