

SUPERVISOR TEST

WASHINGTON COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106406

Date: 11/08/2021

OPERATOR NAME:  
KYLE L KOLAR  
Permit #: 44063  
Operator Agency/Dept: WASHINGTON CO. SO  
Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:55
INTERNAL STANDARD	VERIFIED	18:55
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	18:56
BLANK TEST	0.00	18:57
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	18:57
BLANK TEST	0.00	18:58
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	18:58
BLANK TEST	0.00	18:59

*[Handwritten Signature]* #691

