

SUPERVISOR TEST

 WASHINGTON CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 100284

Date: 11/08/2021

OPERATOR NAME:
 MATTHEW K JACKSON
 Permit #: 40078
 Operator Agency/Dept: WASHINGTON CO. SO
 Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:52
INTERNAL STANDARD	VERIFIED	06:53
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:53
BLANK TEST	0.00	06:54
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:54
BLANK TEST	0.00	06:55
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:55
BLANK TEST	0.00	06:56

