

SOLUTION CHANGE

WHITEHALL VILLAGE  
POLICE DEPT.  
DATAMASTER dmt: 113806

Date: 12/02/2021

OPERATOR NAME:  
ALEXANDER T GILMORE  
Permit #: 47649  
Operator Agency/Dept: WHITEHALL VILL. P.D.  
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:28
INTERNAL STANDARD	VERIFIED	01:28
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	01:28
BLANK TEST	0.00	01:29
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	01:29
BLANK TEST	0.00	01:30
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	01:31
BLANK TEST	0.00	01:31

