

SOLUTION CHANGE

WASHINGTON CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 100284

Date: 12/01/2021

OPERATOR NAME:
MATHEW K JACKSON
Permit #: 40078
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:22
INTERNAL STANDARD	VERIFIED	06:22
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:22
BLANK TEST	0.00	06:23
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:24
BLANK TEST	0.00	06:24
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:25
BLANK TEST	0.00	06:26

