

SUPERVISOR TEST

WASHINGTON CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 100284

Date: 12/15/2021

OPERATOR NAME:
MATTHEW K JACKSON
Permit #: 40078
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:14
INTERNAL STANDARD	VERIFIED	21:14
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	21:14
BLANK TEST	0.00	21:15
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	21:15
BLANK TEST	0.00	21:16
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	21:16
BLANK TEST	0.00	21:17

