

SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 11/27/2021

OPERATOR NAME:
THOMAS J ZOVVISTOSKI
Permit #: 36771
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 21280

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 21:57 |
| INTERNAL STANDARD | VERIFIED | 21:57 |
| REFERENCE STANDARD TEMP | 34.02c | |
| REFERENCE STANDARD | 0.10 | 21:58 |
| BLANK TEST | 0.00 | 21:58 |
| REFERENCE STANDARD TEMP | 34.02c | |
| REFERENCE STANDARD | 0.10 | 21:59 |
| BLANK TEST | 0.00 | 21:59 |
| REFERENCE STANDARD TEMP | 34.02c | |
| REFERENCE STANDARD | 0.10 | 22:00 |
| BLANK TEST | 0.00 | 22:01 |

