

SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 12/13/2021

OPERATOR NAME:
ERNEST R BASSETT
Permit #: 25518
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:00
INTERNAL STANDARD	VERIFIED	15:00
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:00
BLANK TEST	0.00	15:01
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	15:01
BLANK TEST	0.00	15:02
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:03
BLANK TEST	0.00	15:03

