

SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 12/18/2021

OPERATOR NAME:
THOMAS J ZOVISTOSKI
Permit #: 36771
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:36
INTERNAL STANDARD	VERIFIED	16:36
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:36
BLANK TEST	0.00	16:37
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:37
BLANK TEST	0.00	16:38
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:39
BLANK TEST	0.00	16:39

