

SUPERVISOR TEST

HUDSON FALLS
 VILLAGE POLICE DEPT.
 DATAMASTER dmt: 102408

Date: 12/23/2021

OPERATOR NAME:
 LANCE M BARBER
 Permit #: 42841
 Operator Agency/Dept: HUDSON FALLS VIL. PD
 Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:10
INTERNAL STANDARD	VERIFIED	01:10
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:10
BLANK TEST	0.00	01:11
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:12
BLANK TEST	0.00	01:13
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	01:13
BLANK TEST	0.00	01:14

