

SUPERVISOR TEST

HUDSON FALLS  
 VILLAGE POLICE DEPT.  
 DATAMASTER dmt: 102408

Date: 12/08/2021

OPERATOR NAME:  
 LANCE M BARBER  
 Permit #: 42841  
 Operator Agency/Dept: HUDSON FALLS VIL. PD  
 Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:16
INTERNAL STANDARD	VERIFIED	19:16
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	19:16
BLANK TEST	0.00	19:17
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	19:17
BLANK TEST	0.00	19:18
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	19:19
BLANK TEST	0.00	19:19

