

SUPERVISOR TEST

HUDSON FALLS
 VILLAGE POLICE DEPT.
 DATAMASTER dmt: 102408

Date: 12/18/2021

OPERATOR NAME:
 LANCE M BARBER
 Permit #: 42841
 Operator Agency/Dept: HUDSON FALLS VIL. PD
 Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:08
INTERNAL STANDARD	VERIFIED	06:08
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	06:09
BLANK TEST	0.00	06:10
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	06:10
BLANK TEST	0.00	06:11
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	06:11
BLANK TEST	0.00	06:12

