

SUPERVISOR TEST

 HUDSON FALLS
 VILLAGE POLICE DEPT.
 DATAMASTER dmt: 102408

Date: 12/29/2021

OPERATOR NAME:
 DAVID M COSTELLO
 Permit #: 40891
 Operator Agency/Dept: HUDSON FALLS VIL. PD
 Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:11
INTERNAL STANDARD	VERIFIED	15:11
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	15:11
BLANK TEST	0.00	15:12
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	15:12
BLANK TEST	0.00	15:13
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	15:13
BLANK TEST	0.00	15:14

