

SUPERVISOR TEST

WASHINGTON COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106406

Date: 12/27/2021

OPERATOR NAME:  
WILLIAM M GREENE  
Permit #: 48760  
Operator Agency/Dept: WASHINGTON CO. SO  
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:59
INTERNAL STANDARD	VERIFIED	22:00
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	22:00
BLANK TEST	0.00	22:01
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	22:01
BLANK TEST	0.00	22:02
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	22:02
BLANK TEST	0.00	22:03

