

SUPERVISOR TEST

-----  
 WASHINGTON CO.  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 100284  
 -----

Date: 03/23/2021

OPERATOR NAME:  
 ROBERT J SULLIVAN  
 Permit #: 34106  
 Operator Agency/Dept: WASHINGTON CO. SO  
 Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:39
INTERNAL STANDARD	VERIFIED	11:39
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	11:39
BLANK TEST	0.00	11:40
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:41
BLANK TEST	0.00	11:41
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:42
BLANK TEST	0.00	11:43

