

SOLUTION CHANGE

HUDSON FALLS
 VILLAGE POLICE DEPT.
 DATAMASTER dmt: 102408

Date: 01/07/2022

OPERATOR NAME:
 DAVID M COSTELLO
 Permit #: 40891
 Operator Agency/Dept: HUDSON FALLS VIL. PD
 Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:08
INTERNAL STANDARD	VERIFIED	08:08
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	08:09
BLANK TEST	0.00	08:10
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	08:10
BLANK TEST	0.00	08:11
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	08:11
BLANK TEST	0.00	08:12

