

SOLUTION CHANGE

 FORT EDWARD
 VILLAGE POLICE DEPT.
 DATAMASTER dmt: 115708

Date: 12/06/2021

OPERATOR NAME:
 SGT DEAN E WATKINS
 Permit #: 39523
 Operator Agency/Dept: FORT EDWARD VILL. PD
 Reference Standard #: 1100

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:07
INTERNAL STANDARD	VERIFIED	21:07
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:08
BLANK TEST	0.00	21:09
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:09
BLANK TEST	0.00	21:10
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:10
BLANK TEST	0.00	21:11

