

SUPERVISOR TEST

FORT EDWARD
VILLAGE POLICE DEPT.
DATAMASTER dmt: 115708

Date: 12/27/2021

OPERATOR NAME:
SGT DEAN E WATKINS
Permit #: 39523
Operator Agency/Dept: FORT EDWARD VILL. PD
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:43
INTERNAL STANDARD	VERIFIED	20:43
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:43
BLANK TEST	0.00	20:44
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:45
BLANK TEST	0.00	20:46
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:46
BLANK TEST	0.00	20:47

