

WHITEHALL VILLAGE
POLICE DEPT.
DATAMASTER dmt: 113806

Date: 01/08/2022

OPERATOR NAME:
CHRISTOPHER J DAVIDSEN
Permit #: 40077
Operator Agency/Dept: WHITEHALL VILL. P.D.
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:06
INTERNAL STANDARD	VERIFIED	02:06
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	02:07
BLANK TEST	0.00	02:08
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	02:08
BLANK TEST	0.00	02:09
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	02:09
BLANK TEST	0.00	02:10

