

SUPERVISOR TEST

CAMBRIDGE VILLAGE  
POLICE DEPT  
DATAMASTER dmt: 126706

Date: 01/10/2022

OPERATOR NAME:  
ROBERT M DANKO  
Permit #: 39105  
Operator Agency/Dept: CAMBRIDGE VILL. PD  
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:08
INTERNAL STANDARD	VERIFIED	07:09
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	07:09
BLANK TEST	0.00	07:10
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	07:10
BLANK TEST	0.00	07:11
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:11
BLANK TEST	0.00	07:12

