

SUPERVISOR TEST

-----  
GRANVILLE VILLAGE  
POLICE DEPT  
DATAMASTER dmt: 113306  
-----

Date: 01/11/2022

OPERATOR NAME:  
ERNEST R BASSETT  
Permit #: 25518  
Operator Agency/Dept: GRANVILLE VILLAGE PD  
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:53
INTERNAL STANDARD	VERIFIED	15:53
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	15:53
BLANK TEST	0.00	15:54
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	15:54
BLANK TEST	0.00	15:55
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	15:55
BLANK TEST	0.00	15:56

