

SOLUTION CHANGE

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 01/08/2022

OPERATOR NAME:
ROBERT J SULLIVAN
Permit #: 34106
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:34
INTERNAL STANDARD	VERIFIED	14:35
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	14:35
BLANK TEST	0.00	14:36
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	14:36
BLANK TEST	0.00	14:37
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	14:37
BLANK TEST	0.00	14:38

