

SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 01/24/2022

OPERATOR NAME:
THOMAS J ZOISTOSKI
Permit #: 36771
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:56
INTERNAL STANDARD	VERIFIED	03:56
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	03:57
BLANK TEST	0.00	03:57
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	03:58
BLANK TEST	0.00	03:58
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	03:59
BLANK TEST	0.00	04:00

