

SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 01/16/2022

OPERATOR NAME:
CORBET W SULLIVAN
Permit #: 45898
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:26
INTERNAL STANDARD	VERIFIED	08:26
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:26
BLANK TEST	0.00	08:27
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:28
BLANK TEST	0.00	08:28
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:29
BLANK TEST	0.00	08:30

CS
#623

