

SUPERVISOR TEST

 WASHINGTON CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 100284

Date: 01/18/2022

OPERATOR NAME:
 ROBERT J SULLIVAN
 Permit #: 34106
 Operator Agency/Dept: WASHINGTON CO. SO
 Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:55
INTERNAL STANDARD	VERIFIED	06:55
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:56
BLANK TEST	0.00	06:57
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:57
BLANK TEST	0.00	06:58
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:58
BLANK TEST	0.00	06:59

