

SUPERVISOR TEST

HUDSON FALLS  
VILLAGE POLICE DEPT.  
DATAMASTER dmt: 102408

Date: 01/14/2022

OPERATOR NAME:  
LANCE M BARBER  
Permit #: 42841  
Operator Agency/Dept: HUDSON FALLS VIL. PD  
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:14
INTERNAL STANDARD	VERIFIED	23:14
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	23:14
BLANK TEST	0.00	23:15
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	23:15
BLANK TEST	0.00	23:16
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	23:17
BLANK TEST	0.00	23:18

