

SOLUTION CHANGE

HUDSON FALLS
 VILLAGE POLICE DEPT.
 DATAMASTER dmt: 102408

Date: 02/01/2022

OPERATOR NAME:
 DAVID M COSTELLO
 Permit #: 40891
 Operator Agency/Dept: HUDSON FALLS VIL. PD
 Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:54
INTERNAL STANDARD	VERIFIED	08:54
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	08:55
BLANK TEST	0.00	08:56
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	08:56
BLANK TEST	0.00	08:57
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	08:57
BLANK TEST	0.00	08:58

