

SUPERVISOR TEST

HUDSON FALLS
VILLAGE POLICE DEPT.
DATAMASTER dmt: 102408

Date: 02/08/2022

OPERATOR NAME:
LANCE M BARBER
Permit #: 42841
Operator Agency/Dept: HUDSON FALLS VIL. PD
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:53
INTERNAL STANDARD	VERIFIED	02:53
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	02:53
BLANK TEST	0.00	02:54
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	02:55
BLANK TEST	0.00	02:56
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	02:56
BLANK TEST	0.00	02:57

