

SUPERVISOR TEST

WHITEHALL VILLAGE
POLICE DEPT.
DATAMASTER dmt: 113806

Date: 02/05/2022

OPERATOR NAME:
BRYAN J GRECO
Permit #: 44077
Operator Agency/Dept: WHITEHALL VILL. P.D.
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:31
INTERNAL STANDARD	VERIFIED	19:31
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	19:31
BLANK TEST	0.00	19:32
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	19:33
BLANK TEST	0.00	19:33
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	19:34
BLANK TEST	0.00	19:35

