

SOLUTION CHANGE

-----  
 FORT EDWARD  
 VILLAGE POLICE DEPT.  
 DATAMASTER dmt: 115708  
 -----

Date: 01/03/2022

OPERATOR NAME:  
 SGT DEAN E WATKINS  
 Permit #: 39523  
 Operator Agency/Dept: FORT EDWARD VILL. PD  
 Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:07
INTERNAL STANDARD	VERIFIED	16:07
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:07
BLANK TEST	0.00	16:08
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:09
BLANK TEST	0.00	16:09
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:10
BLANK TEST	0.00	16:11

