

SUPERVISOR TEST

FORT EDWARD
VILLAGE POLICE DEPT.
DATAMASTER dmt: 115708

Date: 01/25/2022

OPERATOR NAME:
CORBET SULLIVAN
Permit #: 45898
Operator Agency/Dept: FORT EDWARD VILL. PD
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:12
INTERNAL STANDARD	VERIFIED	11:13
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	11:13
BLANK TEST	0.00	11:14
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	11:14
BLANK TEST	0.00	11:15
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	11:15
BLANK TEST	0.00	11:16

CS
#623

