

SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 01/31/2022

OPERATOR NAME:
THOMAS J ZOVISTOSKI
Permit #: 36771
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:08
INTERNAL STANDARD	VERIFIED	06:08
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	06:09
BLANK TEST	0.00	06:09
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	06:10
BLANK TEST	0.00	06:11
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	06:11
BLANK TEST	0.00	06:12

