

SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 01/24/2022

OPERATOR NAME:
JASON R DIAMOND
Permit #: 37496
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:45
INTERNAL STANDARD	VERIFIED	04:45
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	04:45
BLANK TEST	0.00	04:46
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	04:47
BLANK TEST	0.00	04:48
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	04:48
BLANK TEST	0.00	04:49

