

SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 01/31/2022

OPERATOR NAME:
KYLE L KOLAR
Permit #: 44063
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:27
INTERNAL STANDARD	VERIFIED	19:27
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	19:28
BLANK TEST	0.00	19:28
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	19:29
BLANK TEST	0.00	19:30
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	19:30
BLANK TEST	0.00	19:31

[Handwritten signature] #691

