

SOLUTION CHANGE

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 02/01/2022

OPERATOR NAME:
CHRISTOPHER J MURRAY
Permit #: 46226
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:37
INTERNAL STANDARD	VERIFIED	09:38
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:38
BLANK TEST	0.00	09:39
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:39
BLANK TEST	0.00	09:40
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:40
BLANK TEST	0.00	09:41

