

SUPERVISOR TEST

WASHINGTON COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106406

Date: 02/10/2022

OPERATOR NAME:  
ROBERT J SULLIVAN  
Permit #: 34106  
Operator Agency/Dept: WASHINGTON CO. SO  
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:40
INTERNAL STANDARD	VERIFIED	12:41
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	12:41
BLANK TEST	0.00	12:42
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	12:42
BLANK TEST	0.00	12:43
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	12:43
BLANK TEST	0.00	12:44

