

SUPERVISOR TEST

WASHINGTON CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 100284

Date: 01/24/2022

OPERATOR NAME:  
ERIC C WRIGHT  
Permit #: 49047  
Operator Agency/Dept: WASHINGTON CO. SO  
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:18
INTERNAL STANDARD	VERIFIED	15:18
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:19
BLANK TEST	0.00	15:19
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:20
BLANK TEST	0.00	15:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:21
BLANK TEST	0.00	15:22

