

SUPERVISOR TEST

WASHINGTON CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 100284

Date: 01/31/2022

OPERATOR NAME:  
D TODD RICHE  
Permit #: 33411  
Operator Agency/Dept: WASHINGTON CO. SO  
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:46
INTERNAL STANDARD	VERIFIED	15:46
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:46
BLANK TEST	0.00	15:47
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:47
BLANK TEST	0.00	15:48
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:48
BLANK TEST	0.00	15:49

