

SOLUTION CHANGE

WASHINGTON CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 100284

Date: 02/07/2022

OPERATOR NAME:
ROBERT J SULLIVAN
Permit #: 34106
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:44
INTERNAL STANDARD	VERIFIED	16:44
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:44
BLANK TEST	0.00	16:45
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:45
BLANK TEST	0.00	16:46
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:46
BLANK TEST	0.00	16:47

